



Form GSOP 1-PIN (04/98)

STATE OF CALIFORNIA
Department of General Services - Office of Procurement

PURCHASE ORDER

Purchase Order No.	Rev.	Date
61955		3/ 4/2008

Supplier No.	Solicitation No.	Delivery Date	FOB Point	Invoice Terms
740087	56618	30 Days ARO	Destination	

DVBE TECH INCORPORATED
5060 SUNRISE BLVD
SUITE A-4
FAIR OAKS, CA 95628
Attn: MARK ECKERT

Phone: 916-960-0116

S	DEPT OF WATER RESOURCES	C	WATER RESOURCES	A-36
h	CENTRAL WAREHOUSE	h	ATTN: ACCOUNTS PAYABLE	
i	1721 - 13TH STREET	r	P.O. BOX 942836	
P	SACRAMENTO, CA 95814	e	SACRAMENTO CA 94236-001	

Agency Billing	Agency Purchase Estimate	Purchase Estimate	Revision
81000	6000008107	66602	3

Agency Contact	Phone	Date Received
MIKE E. BINGAMAN	916-654-2522	

Item No.	Quantity	Unit	Commodity Code	Description	Unit Price	Extension
----------	----------	------	----------------	-------------	------------	-----------

THE GENERAL PROVISIONS FOR NON-IT COMMODITIES ARE HEREBY INCORPORATED BY REFERENCE. THESE GENERAL PROVISIONS CAN BE OBTAINED BY PHONING (916) 375-4400 OR BY ACCESSING OUR WEBSITE AT:

www.documents.dgs.ca.gov/pd/modellang/GPnonIT0407.pdf

THE FOLLOWING INFORMATION IS PROVIDED FOR AGENCY USE ONLY:

PRIME CONTRACTOR: SB/DVBE

63	EA	6655-665-0100-2	ACCELEROGRAPH	5,110.0000	321,930.00
----	----	-----------------	---------------	------------	------------

FREE FIELD INTEGRATED TRI-AXIAL ACCELEROGRAPH PER BID SPECIFICATION 6655-07BS-003 OF SIX (6) PAGES DATED JULY 24, 2007.

Brand: REEtek
Model: 130-SMA

Total Value: 321,930.00

F.O.B. DESTINATION REQUIREMENT:

For the purpose of this award only F.O.B. Destination will be accepted.

ONE YEAR MINIMUM WARRANTY AND/OR MANUFACTURER'S WARRANTY REQUIREMENT:

Equipment shall be fully warranted to be free from defects in material and workmanship for a minimum warranty period of one year from date of acceptance by the State or include the standard manufacturer's warranty if longer. Vendor shall bear all material, labor and transportation costs for repair of defects and failures occurring within the warranty period.

Sales and/or use tax to be extra unless noted above

Buyer Shannon Keller SHANNON KELLER	Phone 916-375-4606	BOC Number 14381
--	------------------------------	----------------------------

PURCHASE ORDER CONTINUATION

Form GSOP 2-PIN (04/98)

Page 2 (Last)

<i>Purchase Order No.</i>	<i>Revision</i>	<i>Date</i>	<i>Supplier No.</i>	<i>Supplier Name</i>
61955		3/4/2008	740087	DVBE TECH INCORPORATED

<i>Item No.</i>	<i>Quantity</i>	<i>Unit</i>	<i>Commodity Code</i>	<i>Description</i>	<i>Unit Price</i>	<i>Extension</i>
<p><u>TRAINING REQUIREMENT:</u></p> <p>Training in maintenance and basic operation of the free field integrated tri-axel accelerograph is to be provided at a agency designated location.</p> <p><u>OPERATOR'S MANUAL REQUIREMENT:</u></p> <p>Each unit must come with an operator's manual. All manuals to be in English.</p> <p><u>SALES TAX:</u></p> <p>Sales tax is not to be included on the bid or in the bid pricing. If awarded this bid, sales tax should be added at the time of invoicing. The sales tax rate applied should be based on the rate of the area the product is to be delivered to.</p> <p><u>CHANGE ORDERS:</u></p> <p>Any Purchase Order resulting from this bid may be amended, modified, or terminated at any time by mutual agreement of the parties in writing. Change orders amending, modifying or terminating the Purchase Order, including any modifications of the compensation payable, may be issued only by the State Procurement Officer. All such change orders shall be in writing and issued only upon written concurrence of the supplier. Termination, as that term is used in this section, does not include termination for default of the supplier.</p> <p>This Purchase order has been registered into the state contact and procurement registration system (https://www.scprs.dgs.ca.gov/). The registration number is 38600308290372.</p>						



1 SCOPE

This document establishes requirements and specifications for an integrated digital accelerograph with solid-state memory and internal triaxial accelerometers that;

- 1.1 can function normally in extreme environmental elements of humidity, dust and temperature.
- 1.2 is a standard production model used by leading networks for recording strong motion.
- 1.3 will be utilized by State of California, Department of Water Resources

2 APPLICABLE SPECIFICATIONS / STANDARDS / CODES

2.1 Specifications and standards referenced in this document in effect on the opening of the invitation for bid, form a part of this specification.

2.2 Upon written request, bidder shall submit certification that units and accessories furnished to meet specifications of this document comply with requirements of following standards/codes;

- 2.2.1 FCC § 15 Class A for industrial/commercial electrical equipment.
- 2.2.2 Underwriters Laboratories (UL) standard 50 for electrical equipment safety enclosures.
- 2.2.3 UL standard 508 for Industrial Control Equipment.
- 2.2.4 UL standard 94-V2 for Flammability of Plastic Materials for internal Parts.
- 2.2.5 National Fire Protection Association (NFPA) code 70 for electrical conductors, equipment, signaling and communications conductors and raceways.
- 2.2.6 Ingress Protection (IP) standard 67 for watertight integrity of electrical enclosures.
- 2.2.7 National Electrical Manufacturers Association (NEMA) Type 6 for electrical enclosures.
- 2.2.8 NEMA Type 4 for system enclosures.
- 2.2.9 Electro Static Discharge Association (ESDA) standard S541-2003 for electrical components and enclosures.

3 REQUIREMENTS

3.1 GENERAL OPERATION;

- 3.1.1 Recorder shall remain in a standby condition, and start recording when actuated by strong shaking or manually activated for test purposes. After actuation, it shall stop recording after a set time or after shaking stops, and return to the standby condition, ready to record another earthquake without servicing or attention. Recording shall stop without overwriting previous events when memory becomes full.
- 3.1.2 Unit furnished shall have 3 or more channels capability.
- 3.1.3 Unit shall be battery powered with solid-state memory and internal triaxial Accelerometers.
- 3.1.4 Shall be 18-bit or higher AD conversion, with appropriate anti-aliasing filtering
- 3.1.5 Shall be able to be installed and field-calibrated with a minimum of adjustment.
- 3.1.6 Shall include integrated internal triaxial Accelerometers
- 3.1.7 Record accurately strong shaking (as high as 4g) without data loss or damage.

3.2 SYSTEM CHARACTERISTICS;

- 3.2.1 The noise level of the system (including Accelerometer, Recorder and interconnection) in the absence of shaking shall be a maximum of 0.030 mg (i.e., one count of an 18-bit system with 4g maximum acceleration).
- 3.2.2 Frequency response of the complete transducer-Recorder system shall be flat to within 1% from 0 to 40 Hz, and full documentation shall be provided to allow for accurate instrument correction of the data recorded. Any effects on the recorded data from signals or noise at frequencies higher than 100 Hz shall be <0.030 mg.
- 3.2.3 Recorder and Accelerometer shall reject spurious electronic and electromagnetic noise and interference from radio frequencies (RFI) as well as AC & other frequencies.
- 3.2.4 Recorder, Modem and Accelerometer shall also include effective protection against power and phone line surges and electrostatic discharge (ESD) effects.

3.3 ACCELEROMETERS;

Accelerometers are to be force-balance type meeting the following specifications;

- 3.3.1 **Operating Acceleration Range:** $\pm 4g$; full-scale output voltage of Accelerometer shall match the full-scale input of the Recorder.
- 3.3.2 **Accelerometer Frequency Response (Bandwidth):** Natural frequency shall be >45 Hz.
- 3.3.3 **Linearity:** $\pm 1\%$ or $1000\mu g/g^2$ error from negative to positive full scale.
- 3.3.4 **Accelerometer Noise:** <0.03 mg from 0 to 100 Hz.
- 3.3.5 **Cross-axis Sensitivity:** $\leq 1\%$ g/g.
- 3.3.6 **Temperature Effects:** $<2\%$ over the required operating temperature range of 0 to $150^\circ F$ (-20 to $70^\circ C$). Zero point thermal drift shall be $<500\mu g/^\circ C$.
- 3.3.7 **Function Testing:** Accelerometers shall produce calibration voltages during static tilt tests on all channels to allow verification of sensitivity constants. Accelerometer operation shall be verifiable via signal offset in response to signal command issued from the Recorder.
- 3.3.8 **Accelerometer Zero Output (Zero Adjust):** Adjustable to 0, $\pm 10mg$ after field installation to either a vertical or horizontal surface. Accelerometer zero value shall be stable over time, without steps or other changes during recording. Accelerometer zero adjustment mechanism, if present, shall lock securely so that the Accelerometer output resumes the pre-event zero level after earthquake shaking. If auto-zeroing of DC level is possible, it shall have the ability to be deactivated.

3.4 RECORDER;

The accelerograph shall have unity gain in fixed steps, and meet the following specifications;

- 3.4.1 **Sampling:** Recorded data sampling rate shall be 200 samples per second per channel. Anti-alias filtering of the analog signal from the Accelerometers or extraneous sources shall meet the System Characteristics noise specification (see 3.2). Channel-to-channel sampling to be synchronized to within 0.2 msec.

- 3.4.2 **Recorder Noise:** During normal operation, and when the Recorder transitions from the standby state to recording operation, the noise level shall not exceed the System Characteristics noise specification (see 3.2).
- 3.4.3 **Temperature Effects:** <2% over the required operating temperature range of 0 to 150°F (-20 to 70°C).
- 3.4.4 **GPS Timing:** Recorder shall record sample time information from the internal or external GPS receiver in the event header and in each file data block. Accelerograph shall also include an internal clock that is accurate to within 1 minute per month in the absence of GPS. Duty cycling of the GPS receiver shall be settable from 0% to 100%. Time and amount of last GPS update, GPS status information, and time-averaged location information (latitude, longitude and elevation) shall be available from the event header and via a status command. GPS receiver shall update internal clock when drift is greater than 0.5 msec. GPS antenna and receiver shall be locatable up to 1000' (300 m) from the Recorder without performance problems.
- 3.4.5 **Event Run Time:** Pre-event run time shall be adjustable from 1 to 30 seconds. Post-event hold-on time shall be adjustable from 10 to 60 seconds.
- 3.4.6 **Triggering:** Declaration of a triggered condition to depend on the acceleration (within the 0.1 - 12 Hz frequency band) and channel trigger voting. When the acceleration on a specified channel exceeds the trigger threshold level specified for that channel, that channel's specified number of votes is added to the total votes for trigger. When the total number of votes reaches the minimum specified for triggering, a trigger condition is declared, and recording is to begin at the start of the next second. Once a triggered condition is declared, recording shall continue until after the detriquer condition is met, dependent on the continuing motion and the channel detriquer voting. When the acceleration on a channel no longer exceeds the detriquer threshold level specified for that channel, continuously for the number of seconds specified by the post-event run time, that channel's specified number of votes is added to the total votes for detriquer, for as long as that condition continues. When the total number of votes reaches the specified minimum for detriquer, a detriquer condition is declared, and recording shall stop at the start of the next second. For all channel signals, trigger and detriquer threshold levels shall be independently adjustable from 0.1 mg to full-scale. The number of votes for each channel shall be settable from 0 through 5 for both trigger and detriquer. The total number of votes required to trigger or detriquer the Recorder shall be independently settable from 0 through 9.
- 3.4.7 **Function Tests:** Under user control, Recorder shall generate test signal commands to the Accelerometers to allow verifying Accelerometer operation from the recorded data. Recorder shall accommodate and record tilt tests on all 3 axes to allow calibration of the sensitivity constants of the Accelerometers.
- 3.4.8 **Diagnostics / State of Health Monitoring:** Upon power up, and under user control at other times, Recorder shall perform internal self-checks and diagnostics to verify proper operation. Under user control, Recorder shall also provide, without generating an event, amount of available memory, firmware version number, timing, battery voltages, recording parameters and

Accelerometer output levels (ambient and with offset signal applied). All Recorder configuration and operating parameters shall be written into a retrievable text file with a single command.

- 3.4.9 **Data Storage:** ≥64Mb of storage memory, storable on standard memory cards (e.g., PCMCIA flash memory card) readable by laptop PCs (MS[®] Operating System). Files on memory cards shall be individually deletable as well as globally deletable.
- 3.4.10 **Batteries:** Entire system shall be powered by sealed internal and/or external 12-volt rechargeable DC batteries. Lithium batteries (if any) shall have a sealed stainless steel case. All batteries shall have the capability to be replaced without soldering.
- 3.4.11 **Battery Charging:** Recorder shall include AC charging circuitry for maintaining the charge of the internal and/or external main batteries. Charger must be able to charge a completely discharged battery, and charging voltage and current shall be held within battery manufacturer guidelines.
- 3.4.12 **Recovery after Loss of Power:** In the event of main battery failure, Recorder shall automatically return to normal operation (i.e., able to trigger and record data) once power is restored, without affecting retrievability of previously recorded data. In the event of long-term power loss, recorded data shall be preserved in memory for at least 30 days without affecting data retrievability.
- 3.4.13 **Power Consumption:** Total power consumption at standby, to facilitate solar-powered deployment, shall be ≤55 watt-hrs/day (at 12 vDC nominal), including the power for the sensors, the power for the GPS timing receiver with 5% duty cycling, and the power for the modem (see 3.5.2) in the absence of calls. When installed, internal battery(ies) shall provide for autonomous operation of at least 36 hours. Internal charger shall be capable of charging sufficient external batteries for 4 days autonomous operations.
- 3.4.14 **Accelerograph Housing:** Compact (≤1³), total weight ≤ 40lbs (18Kg) when fully equipped with internal sensors and batteries, and includes a hole for locking that accommodates a standard ⁵/₁₆" shackle padlock. Housing shall be suitable for outdoor deployment, made of engineering plastic, galvanized or stainless steel, crush resistant (able to withstand 30lb weight drop from 3'), resistant to EMI/RFI, NEMA 6/IP67 rated and include a rigid mounting base. Accelerometers shall be securely mounted orthogonally in the accelerograph housing, perpendicular and parallel to the case sides, within 1^o.
- 3.4.15 **Deployment:** Accelerograph shall be easily secured (with at least one mounting bolt accessible only from inside) to horizontal or vertical surfaces; mounting base shall have three protruding points of contact to facilitate leveling and mounting on irregular surfaces without deformation of mounting base.
- 3.4.16 **Electrical Connectors:** All external electrical connectors shall be of the sealed "MS" type. Accelerograph shall allow input (with an optional connector) from three external sensors ≤2000' away, in lieu of the internal sensors. Any externally accessible switches shall require a key to operate.

3.5 COMMUNICATION WITH RECORDER;

- 3.5.1 Remote Communication:** Remote interrogation of the system state-of-health shall be available via the included 56 kbps or faster telephone modem (or optional cellular modem or TCP/IP adapter) without retrieving or altering the data in memory. The state-of-health information shall include battery voltages, battery charging status, GPS system status, amount of memory used, event count, and accelerogram parameters (including peak values, event length, time of trigger and channels causing trigger). The remote access shall allow performing all functions that the local on-site connection can perform (including data retrieval, upgrading the firmware, and functional and diagnostic tests) and shall include protection against unauthorized access. All interrogation and control shall be by means of ASCII-character commands from a PC in terminal mode.
- 3.5.2 Modem Power Control:** Modem (internal or external) shall normally be not powered, or in the absence of calls, use <3 watt-hrs/day at 12vDC. Modem, when active, shall use <225mA at 12vDC nominal. The Recorder shall, upon triggering, power the modem and initiate call-out within 5 seconds. For incoming calls, a ring-detect provision shall turn on the modem and answer the call. Modem control shall include ≥ 5 specifiable windows of time during which the modem and optional external cellular phone are powered. Modem and communication equipment shall remain powered for the duration of the call.
- 3.5.3 Triggered Communication:** Upon being triggered, Recorder shall auto-dial a remote computer, initiating dialing within <5 seconds. Recorder shall identify the name of the event file causing it to call-out. If the communication is unsuccessful, the unit shall re-attempt to communicate after successive time intervals doubling in length, starting at 15 seconds and increasing to 3 days, and maintain a retrievable log of the communication attempts. Recorder shall declare communication success upon receipt of a specific command from the remote computer.
- 3.5.4 Alarm Communication:** Recorder shall call a remote computer (see 3.5.3) upon detecting selectable equipment malfunctions (including at least low battery level and loss of AC) and indicate the reason for the call. Recorder shall also call out at a specifiable time or when user specifiable channels exceed specifiable "alarm" threshold levels.
- 3.5.5 Triggered Data Retrieval:** Recorded data shall be retrievable locally or remotely via modem, using a desktop or laptop PC and communication software using terminal mode and standard file transfer protocols (Xmodem, Ymodem, Zmodem), allowing data from the Recorder to be transferred rapidly to standard PC media. During data retrieval and interrogation, the system shall remain operational and record any events that occur. Peak acceleration values shall be updated during the event and be available for retrieval while the event is still being recorded. Data retrieval from memory shall yield an exact copy of the file, and not cause stored data to be erased. Vendor shall provide specification, operating system and interface software for the MS® Operating System Recovery computer.
- 3.5.6 Data Transfer:** Rate Data transfer rate (per three channels) shall be such that 1 minute of three-channel data is transferred in <1 minute (that is, on a per-channel basis, a transfer time of <20 seconds/channel for 60 sec. of data). Compression algorithm (if any) shall have a data-reconstruction error of <1 LSB from 0Hz to the Nyquist frequency.

- 3.5.7 Data Handling Software:** Data review and retrieval software shall be provided which operates on a MS[®] Operating System PC and provides for field review of data as well as output in an event file with well-documented format. Software shall include a stand-alone program to display / print event header information in summary form. Full documentation of the header and data layout in the raw event file shall be provided. File conversion software shall be provided (with documented source code) to convert the event file from native format to COSMOS Volume 0 Strong Motion Data Format, v 1.20 (<http://www.cosmos-eq.org>).
- 3.5.8 Continuous Data Transmission:** Simultaneously with other operations, instrument shall be capable of continuous transmission of a digital data stream for full duplex real-time communication. Unit shall use a fully documented protocol compatible with modern communications (Frame Relay, DSL, TCP/IP, etc) for input into a digital seismic monitoring system (e.g., Earthworm, etc.).

3.6 SYSTEM ITEM CONFIGURATION;

3-Channel Recorder, ≥ 18 bit with internal triaxial accelerometer with Force-feedback of $\pm 4g$, 64Mb removable memory, GPS receiver with antenna and 25m coax cable, internal 56Kb modem, interconnect fully populated. Bid for a quantity of 63.