



Form GSOP 1-PIN (04/98)

STATE OF CALIFORNIA
 Department of General Services - Office of Procurement
PURCHASE ORDER

Purchase Order No. Rev. Date
62142 6/30/2008

Supplier No. 808590	Solicitation No. 57025	Delivery Date 30 Days ARO	FOB Point Destination	Invoice Terms
S CA DEPT OF PUBLIC HEALTH h CHILDHOOD LEAD POISONING i 850 MARINA BAY PKWY B 3 p RICHMOND, CA 94804-6403 Attn: C. KINSMAN (5100 620-5600)		C CA DEPT PUBLIC HEALTH h CPSS PURCHASING SERV UNIT a PO BOX 997377, MS 1802 t SACRAMENTO, CA 95899-7377 r o e		
Agency Billing 83718	Agency Purchase Estimate 22331-07-5370	Purchase Estimate 67138	Revision 0	
Agency Contact CHRISTINE KINSMAN		Phone 510-620-5600	Date Received	

THERMO NITON ANALYZERS LLC
 900 MIDDLESEX TURNPIKE
 BLDG #8
 BILLERICA, MA 01821
 Attn: ALICIA FLOOD

Phone: 877-255-6943

Item No.	Quantity	Unit	Commodity Code	Description	Unit Price	Extension
<p>THE GENERAL PROVISIONS FOR NON-IT COMMODITIES ARE HEREBY INCORPORATED BY REFERENCE. THESE GENERAL PROVISIONS CAN BE OBTAINED BY PHONING (916) 375-4400 OR BY ACCESSING OUR WEBSITE AT: www.documents.dgs.ca.gov/pd/modellang/GPnonIT0407.pdf</p> <p>THE FOLLOWING INFORMATION IS PROVIDED FOR AGENCY USE ONLY:</p> <p>PRIME CONTRACTOR: NS DISCAL YEAR: 2007/2008</p>						
1	16	EA	6630-999-9888-0	ANALYZER (AS DESCRIBED) LEAD ANALYZER (AS DESCRIBED) IN ACCORDANCE WITH SPECIFICATION # 6630-234 OF ONE (1) PAGE, DATED JULY 1, 2008.	20,000.0000	320,000.00
				Brand: THERMO NITON ANALYZER		
				Model: XLP 303A		
				<u>PO Miscellaneous Charges and Discounts</u>	<u>Dollar Value</u>	
				DELIVERY CHARGE		3,120.00
				BATCH ADJUSTMENT (DISCOUNT)	-	16,000.00
				TRADE DISCOUNT IN DOLLARS	-	24,000.00
Total Value:						283,120.00
ATTACHMENTS						
THE FOLLOWING ATTACHED DOCUMENTS ARE PART OF THIS PURCHASE ORDER. SPECIFICATIONS-# 6630-234 OF ONE (1) PAGE, DATED JULY 1, 2008.						
F.O.B. DESTINATION						
For the purpose of this order only F.O.B. Destination will be accepted.						

Sales and/or use tax to be extra unless noted above

Buyer LONNIE WILLIAMS	Phone 916-375-4586	BOC Number
-------------------------------------	------------------------------	-------------------

STATE OF CALIFORNIA

Department of General Services - Office of Procurement

PURCHASE ORDER CONTINUATION

Form GSOP 2-PIN (04/98)

Page 2 (Last)

<i>Purchase Order No.</i> 62142	<i>Revision</i>	<i>Date</i> 6/30/2008	<i>Supplier No.</i> 808590	<i>Supplier Name</i> THERMO NITON ANALYZERS LLC
---	-----------------	--------------------------	-------------------------------	--

<i>Item No.</i>	<i>Quantity</i>	<i>Unit</i>	<i>Commodity Code</i>	<i>Description</i>	<i>Unit Price</i>	<i>Extension</i>
<p><u>DELIVERY INFORMATION</u></p> <p>Delivery date will be thirty (30) DAYS ARO.</p> <p>Delivery will be made during normal business hours, Monday - Friday, from 8:00 am to 3:30 pm, except State holidays.</p> <p>Deliver to: California Department of Public Health Childhood Lead Poisoning 850 Marnia Bay Parkway B3 Richmond, CA 94804-6403 Attn: Julet Rafol</p> <p>Please contact agency seven (7) days prior to delivery to make arrangements for delivery.</p> <p>Contact: Julet Rafol Phone: (510) 620-5616</p> <p><u>MANUALS, TOOLS, TRAINING, WARRANTY</u></p> <p>MANUALS / TOOLS:</p> <p>Each system shall come with one operational manual, service manual and parts list (if applicable). All manuals shall be in the English language.</p> <p>If there are any special tools required for basic maintenance, one set shall be provided.</p> <p>TRAINING:</p> <p>Training shall be provide for a minimum of thirty (30) users on the use and maintenance of the instrument. Training material shall be provided for all 30 users.</p> <p>WARRANTY:</p> <p>Equipment shall operate satisfactorily and have a minimum warranty period of one year from date of acceptance to the State. Supplier shall bear all material, labor and transportation cost for repair of defects and failures occurring with in the warranty period.</p> <p><u>AWARD DATE</u></p> <p>This purchase order is being awarded on September 8, 2008 pursuant to Government Code Section 13332.17. Any encumbrances made pursuant to this purchase order shall be construed to have been made on the last day of the preceding fiscal year.</p> <p><u>SCPRS REGISTRATION</u></p> <p>This Purchase order has been registered into the state contact and procurement registration system (https://www.scprs.dgs.ca.gov/). The registration number is 42650908331892.</p> <p><u>CHANGE ORDERS</u></p> <p>Purchase Order may be amended, modified, or terminated at any time by mutual agreement of the parties in writing. Change orders amending, modifying or terminating the Purchase Order, including any modifications of the compensation payable, may be issued only by the State Procurement Officer. all such change orders shall be in writing and issued only upon written concurrence of the supplier. Termination, as that term is used in this section, does not include termination for default of the supplier.</p>						

**1. Scope**

The requirement herein describes a handheld XRF lead-based paint analyzer for the Department of Public Health, Childhood Lead Poisoning. The System includes an XRF lead-based paint analyzer, rechargeable batteries, carrying case, User manual, software/hardware, and test stand.

2. Applicable Specifications

Specifications and standards referenced in this document in effect on the opening of the Invitation for Bid form a part of this specification where referenced.

3. Requirements

- 3.1 Shall be an X-ray Fluorescence (XRF) type analyzer.
- 3.2 Shall be a portable unit that weights less than 4 pounds.
- 3.3 The offered analyzer shall have a Performance Characteristic Sheet (PCS) issued by Midwest Research Institute on behalf of the U.S. Department of housing (HUD) and the U.S. Environmental Protection agency (EPA). The PCS shall be on the HUD website.
- 3.4 The analyzer shall be capable of dust wipe and soil sample testing.
- 3.5 National Institute of Standards and Technology (NIST) Standard Reference Material (SRM) calibration standards for lead-based paint and calibration standards for lead in dust wipe and soil samples must be provided.
- 3.6 The Systems shall be equipped with, but not limited to, the following:
 - 3.6.1 Portable Test Stand for soil and dust wipe testing.
 - 3.6.2 Two rechargeable batteries and charger.
 - 3.6.3 Lockable carrying case that secures the analyzer and contents from movement.
 - 3.6.4 User manual
 - 3.6.5 Interfacing software and hardware compatible with a Windows® based computer.
 - Software shall be supplied, including a backup CD, and fully supported by the vendor.
 - Software shall be used to transfer data for report generation.